

Synthetic Megavoltage Cone Beam CT: A Tool for Improved Pacemaker Contouring Accuracy

Andreas Stoya*

Department of Anesthesiology and Intensive Care, University of Leipzig Medical Center, Liebigstrasse 20, 04103 Leipzig, Germany

Introduction

Cardiac pacemakers have revolutionized the treatment of heart rhythm disorders, providing a lifeline for millions of patients worldwide. These devices are implanted to regulate and stabilize the heart's electrical activity, ensuring that it beats at a consistent rate. However, accurate contouring of cardiac pacemakers within the body during radiotherapy procedures can be challenging due to the limitations of imaging techniques. This challenge has led to the development of novel solutions, one of which is the use of synthetic Megavoltage Cone Beam CT (sMV-CBCT) images to enhance contouring accuracy. In this article, we will explore the significance of cardiac pacemaker contouring accuracy, the limitations of traditional imaging methods, and the role of sMV-CBCT images in improving the precision of radiotherapy for patients with cardiac pacemakers. We will discuss the principles behind sMV-CBCT, its potential applications, and the advantages it offers in cardiac pacemaker contouring [1,2].

Description

Radiotherapy involves the targeted delivery of high-energy radiation to specific areas of the body to treat cancer. Ensuring the cardiac pacemaker is accurately contoured helps minimize radiation exposure to the device and surrounding tissues, reducing the risk of device malfunction and potential harm to the patient. Accurate contouring enables clinicians to plan and deliver radiotherapy with greater precision. This can improve the effectiveness of the treatment by maximizing the dose to the tumor while sparing healthy tissues, including the heart and surrounding organs. Knowing the exact location and geometry of the pacemaker within the patient's body is essential for developing a comprehensive treatment plan. This plan determines the optimal radiation beam angles and intensity to achieve therapeutic goals while minimizing side effects. Patients with cardiac pacemakers already face health challenges, and radiotherapy should not exacerbate these issues. Accurate contouring helps maintain a patient's quality of life by reducing the potential for complications or the need for additional medical interventions. Cardiac pacemakers are predominantly composed of metal, which can cause significant artifacts in conventional CT imaging.

Conclusion

Enhancing cardiac pacemaker contouring accuracy in radiotherapy is of paramount importance to ensure patient safety, treatment efficacy, and quality of life. Traditional imaging methods have inherent limitations, which are

***Address for Correspondence:** Andreas Stoya, Department of Anesthesiology and Intensive Care, University of Leipzig Medical Center, Liebigstrasse 20, 04103 Leipzig, Germany; E-mail: andreas05@stoyanov.edu

Copyright: © 2024 Stoya A. This is an open-access article distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Received: 01 December, 2024, Manuscript No. japre-25-160329; **Editor Assigned:** 04 December, 2024, PreQC No. P-160329; **Reviewed:** 14 December, 2024, QC No. Q-160329; **Revised:** 20 December, 2024, Manuscript No. R-160329; **Published:** 28 December 2024, DOI: 10.37421/2684-5997.2024.7.272

addressed by the innovative approach of synthetic Megavoltage Cone Beam CT (sMV-CBCT). This advanced imaging technique reduces metal artifacts, provides detailed 3D visualization, and offers real-time monitoring capabilities, making it an invaluable tool for cardiac pacemaker contouring. The application of sMV-CBCT in cardiac pacemaker contouring brings significant advantages, including enhanced accuracy, patient safety, optimized treatment plans, time efficiency, and increased patient satisfaction. As research and technology continue to evolve, sMV-CBCT is poised to become a standard in radiotherapy, contributing to improved outcomes and a higher standard of care for patients with cardiac pacemakers. In the coming years, ongoing research and clinical validation will further refine the implementation of sMV-CBCT, making it an indispensable component of the radiotherapy toolkit for the benefit of both patients and healthcare providers.

References

1. Horvath, Balazs, Benjamin Kloesel, Michael M. Todd and Daniel J. Cole, et al. "The evolution, current value, and future of the american society of anesthesiologists physical status classification system." *Anesthesiology* 135 (2021): 904-919.
2. Riley, Richard D., Joie Ensor, Kym IE Snell and Frank E. Harrell, et al. "Calculating the sample size required for developing a clinical prediction model." *BMJ* 368 (2020).

How to cite this article: Stoya, Andreas. "Synthetic Megavoltage Cone Beam CT: A Tool for Improved Pacemaker Contouring Accuracy." *J Anesthesiol Pain Res* 7 (2024): 272.